



This Pro CO Alarm Analyzer combines carbon monoxide detection, a low pressure alarm, an external high temperature sensor and a low pressure flow meter / regulator. Components are mounted in a water and impact resistant case with a clear cover that is compatible with outdoor and marine environments. Available in two power options.

## FEATURES

- Tank readings through restrictor or in-line monitoring on compressor
- On / off button
- Temperature compensated sensor
- User replaceable battery options
- Low battery warning indicator
- Electrochemical sensor

## SPECIFICATIONS\*

Flow Rate	0.5 - 5 L/min
Resolution	1 ppm
Repeatability	<+5%
Accuracy	±5%
Sensor Type	Electrochemical
Expected Sensor Life	>24 months in normal use from date of manufacture
Range	0-50 ppm CO
Alarms	Two user-programmable audible and visual alarms
Response Time	<50 seconds over complete temperature range
Operating Temperature	14° to 122°F (-10° to 50°C) continuous -4° to 122°F (-22° to 50°C) intermittent
Operating Humidity	Non-condensing: 15-90% continuous, 0-99% intermittent
Storage Temperature	14° to 140°F (-10° to 60°C)
Power	Rechargeable lithium battery or 110/220 V wall plug-in
Dimensions	6 x 6 x 5 in (15 x 15 x 13 cm)
Warranty	12 months from date of purchase

## ADVANTAGES

- Two programmable audible & visual threshold alarms
- Alarm at set point or 10 ppm CO
- Fast response
- Made to test breathing gases\*\*
- Easy to operate, reliable and accurate
- Optional relays for external alarm or compressor control

NOTE: Never expose gas sensors to pressure or you may cause damage and/or false readings. Damaged sensors will not provide accurate gas analysis. Inaccurate gas analysis can lead to serious personal injury or death. Most gas analyzers can be used to analyze a regulated gas sample flow, the contents of a gas cylinder, or the flow from a regulator. The flow rate of gas must equal 0.5-5 L/min. To produce this flow, a Flow Restrictor and Regulator may be required.

\*All specifications are at ambient / sea level, 77°F (25°C) and are subject to change without notice. \*\*Calibration must be confirmed with calibration CO test gas.